



California Medical Device Recall Information



Recall Name

Praxair Recalls Grab 'n Go Vantage Portable Oxygen Cylinder Due to Possible Fires

Recall Date	Product Description	Recalling Firm	Recall Reason
11/16/12	Grab 'n Go Vantage Portable Oxygen Cylinder	Praxair, Inc. Danbury, CT	<i>Possible fires may occur if cylinder is knocked over and subject to significant physical impact</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Grab 'n Go Vantage Portable Oxygen Cylinder Suspect Catalog Numbers Recalled: <ul style="list-style-type: none">• OX-MAEGNGVNTG• WESPRX-9500 Product Label	CA , global	Manufactured and distributed from June 17, 2009 to November 16, 2012

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm333846.htm>